


These instructions are valid for all KENDA Dental Polishers that are used for restorative and preventive work.

Single-patient polishers are labelled with  and must not be used more than once, as multi-use of these polishers can cause unnecessary risk to the patient and staff.

Composition: The polishers consist of synthetic/silicone rubber, pigments and abrasive. The shanks are made of stainless steel or medical plastic.

Indication: KENDA Dental Polishers are used for polishing different materials which are specified on the packaging label.

Contraindications: If a patient is known to be allergic to any of the above components, the polishers should not be used.

Adverse effect: None known to date.

General security advices:



Do not exceed maximum speed



Wear breathing protection



Wear eye protection



Use waterspray minimum 50ml/min

Precautions:

- Products may only be used in the specified field of application.
- Check the safe connection of the polisher before using it and perform a test run.
- Use proper hand piece only and remove any damaged, bent, or no longer central running polishers immediately.
- Avoid excessive pressure which can cause tilting, bending or blocking of the polisher.
- Insufficient watercooling can result in irreversible damage to the tooth and the surrounding tissue.
- Using the wrong disinfectant and/or prolonged disinfection times can soak the polisher or cause discoloration. Therefore, make sure that the effectiveness of the disinfectant is approved and suitable for polishers, do not use disinfectants containing aldehydes and follow the manufacturer's specifications of the product with regard to the concentration and reaction time. Ensure that the process is effective.
- Always keep track of LOT numbers of polishers to ensure traceability.

Storage: A dry, dust-free storage at 2 - 28°C (36 - 82°F) is recommended. Improper storage conditions can shorten the shelf life and cause malfunction of the product.

Expiry date: See imprint on label.

Hygiene instructions: Please refer to back of this page for detailed hygiene instructions.

alb	version shqiptar	jpn	日本語版
ara	النسخة العربية	kor	한국어 버전
bul	българска версия	lav	Latviešu valodā
chi	中國版	lit	Lietuviška versija
cze	Česká verze	mac	македонска верзија
dan	danske version	nor	Norsk versjon
dut	Nederlandse	per	نسخه فارسی
est	Eesti versioon	pol	Polska wersja
fin	Suomalainen versio	por	Versão em português
fre	Version française	rum	Versiunea romana
ger	deutsche Version	rus	Русская версия
gre	Ελληνική έκδοση	slo	slovenská verzia
heb	גרסה עברית	slv	slovenska različica
hrv	hrvatska verzija	spa	Versión en español
hun	magyar változat	swe	Svensk version
ice	Íslensk útgáfa	tur	Türkçe versiyonu
ind	Bahasa Indonesia versi	ukr	Українська версія
ita	Versione italiana		




QR-Code
<http://knd.li/ins>

These hygiene instructions are valid for all KENDA Dental Polishers for chairside use.

WARNINGS:

KENDA Polishers are supplied unsterile and therefore must be treated before use.

Single patient polishers (labelled with ) are not eligible for reuse, but need to be disinfected and sterilised before and only use.

During processing of all polishers, protective clothing, gloves and goggles must be worn at all times.

Use bur bath approved and suitable for polishers only.

Limitation of reprocessing: The product's life is determined by the wear and tear of the polishing instrument. Repeated processing has minimal effect on these instruments.

Point of use: Remove surface contamination with a disposable cloth immediately after using on patient and put in a container.

Containment and transportation: Polishers should be transported immediately in a closed instrument tray to the treatment area. Proceed with cleaning and disinfection immediately.

Cleaning and disinfection: Proceed with either machine (recommended) or manual treatment.

a. Machine treatment:

1. Rinse thoroughly under running water.
2. Place polishers in the washer/disinfector (WD) and make sure they do not touch each other.
3. Add the appropriate cleaning agent (aldehyde-free), according to the information on the product label and the instructions of the WD-manufacturer.
4. Start the desired program: cleaning and disinfection.
5. After termination of the program, take the polishers out of the WD and dry them preferably with clean compressed air.
6. Visual inspection for cleanliness and integrity using a magnifying glass with 6 times magnification. In case of visual contamination, the process must be repeated.
7. If only mechanical cleaning (without proven disinfection) a final disinfection in a steam sterilizer (unpacked) in suitable racks or trays is required.
8. Documented approval of the proper treatment procedure in order to reuse or for storage.
9. A dry, dust-free storage at 2 - 28°C (36 - 82°F) is recommended.

b. Manual treatment:

1. Rinse thoroughly under running water. If necessary, use a nylon brush. NOTE: Use a suitable brush for cleaning blind holes.
2. Cleaning and chemical disinfection with a suitable solution (aldehyde-free) according to manufacturer's instruction. Insert polisher into the solution. Make sure that polishers are free of air bubbles.
3. The reaction time begins with the insertion of the last polisher and is according to the instruction of the disinfectant manufacturer.
4. After contact time is up, rinse polishers with plenty of deionized water.
5. Dry residual moisture preferably with clean compressed air.
6. Visual inspection for cleanliness and integrity using a magnifying glass with 6 times magnification. In case of visual contamination, the process must be repeated.
7. Without a proven disinfection, a final disinfection in a steam steriliser (unpacked) in suitable racks or trays is required.
8. Documented approval of the proper treatment procedure in order to reuse or for storage.
9. A dry, dust-free storage at 2 - 28°C (36 - 82°F) is recommended.

Sterilisation: Moist heat sterilisation with fractional prevacuum. Full cycle, 134°C (273°F) / 3,1 bar for at least 5 minutes, drying time 10 minutes. Do not exceed 135°C (275°F).

IMPORTANT: Use suitable packaging for sterilisation.

Warning: KENDA Dental Polishers are not suitable for sterilisation with hot air or chemiclave.

This procedure was validated by an independent and accredited laboratory.



Additional information: When sterilising multiple instruments in one autoclave cycle ensure that the steriliser's maximum load is not exceeded in accordance with the steriliser's instructions.

Manufacturer contact:

KENDA AG
Kanalstrasse 18
LI-9490 Vaduz, Liechtenstein
<http://www.kenda-dental.com>



Date of preparation: 09/2013; Rev. 2

The instructions provided above have been validated by the medical device manufacturer as being capable of preparing KENDA Dental Polishers for reuse.

The material has been developed solely for use in dentistry. Processing should be carried out strictly according to the Hygiene Instructions. Liability cannot be accepted for damages resulting from failure to observe the instructions or the stipulated area of application. The user is responsible for testing the products for their suitability and use for any purpose not explicitly stated in the instructions. Descriptions and data constitute no warranty of attributes and are not binding.

Keep out of reach of children.

For use in dentistry only.

Rx only. Caution: U.S. federal law restricts this device to sale by or on the order of a licensed dentist.

